



- MSM-****-04-T-K2-OP1: II 1G/1D, authorized for Ex zones 0, 1, 2, 20, 21, 22
- MSM-****-04-T-K2-OP2: II 2G/2D, authorized for Ex zones 1, 2, 21, 22
- MSM-****-04-T-K2: Applicable in Non-Hazardous Locations

Ex op is IIC T4 Ga / Ex op is IIIB T135°C Da
or
Ex op is IIB T4 Gb / Ex op is IIIB T135°C Db



II 1G IIC T4 Ga / II 1D IIIC T135°C Da
or
II 2G IIB T4 Gb / II 2D IIIB T135°C Db

Technical data	Type	MSM-****-04-T-K2-OP1	MSM-****-04-T-K2-OP2	MSM-****-04-T-K2
Standard length and designation		****=Length in mm, from 0200mm up to 7000mm (Overall length)		
Ex Protection, Gas		II 1G IIC T4 Ga	II 2G IIB T4 Gb	none
Ex Protection, Dust		II 1D IIIC T135°C Da	II 2D IIIB T135°C Db	none
For use in Ex Zones		0, 1, 2, 20, 21, 22	1, 2, 21, 22	--
Requirement at connected sensors		Ex op is Ga/Da	Ex op is Gb/Db	none
Maximum optical input power		<=15mW	<=35mW	not limited
Maximum potential radiant intensity		<=5mW/mm²	<=5mW/mm²	not limited
Active fibre optic diameter		4 mm		
Active cross-sectional area		12.6mm²		
Transmission rate, average		50-70%, at 870nm		
Optical aperture		appr. 65°, at 870nm		
Individual fibre diameter		50µm		
Minimum bending radius		50mm (Single bend)		
Operating temperature range T _{amb}		0°C < T _{amb} < +120°C		
Enclosure rating, according to EN 60529		IP 65		
Material, probe tip		Brass		
Material, protection sheath		Brass, chromium plated		
Accessories, included		2 x Shrink-down plastic tubing	--	
Accessories, not included		--		
Options		--		

ATEX/IECEx RELATED MARKINGS

CE 0158 T_{amb} = 0°C < T_{amb} < +120°C

Manufacturer with address

Date of production: Numerals 5 to 8 of the serial number (year / calendar week)

Date of production: Numerals 5 to 8 of the serial number (year/calendar week)

Certifications IECEx & ATEX:

Type: MSM-****-04-T-K2-OP1



II 1G IIC T4 Ga, II 1D IIIC T135°C Da

EC-Certification No: BVS 10 ATEX E 130 X. DEKRA

Type: MSM-****-04-T-K2-OP2






II 2G IIB T4 Gb, II 2D IIIB T135°C Db

IECEx-Certification No: IECEx 14.0108X

(X designation of the certification number: Fibre optics must only be applied with sensors with certificated limited optical power)

Dimensions:

Important: The measurerent head should be mounted in such way that both lwl outputs are mounted in a horizontal position!



Operating Manual / EU - Declaration of Conformity:

Installation prescriptions for Ex hazardous locations

Type MSM-****-04-T-K2-OP1: Applicable in Ex zones 0, 1, 2, 20, 21, 22.

Type MSM-****-04-T-K2-OP2: Only applicable in Ex zones 1, 2, 21, 22.

General regulations for all types:

The maximum rated optical input power must not be exceeded. The local equipotential bonding have to be done by grounding the fixed sensor. It is necessary to take into consideration the valid international and national rules and regulations (EN 60079-14). Other then original manufacturer, additional optical lenses are not allowed in hazardous locations. The fibre optics have to be installed in a manner to avoid tensile stress and frictional heat. If fibre optics and associated sensors are not mounted in the same hazardous location, the change over of the different areas must be realized in accordance with the valid regulations. With the additional shrink-down plastic tubings, only types MSM-****-04-T-K2-OP1, a required change over can be realized.

Function

The fibre optics series MSM are designed for the arrangements in hazardous locations and for high ambient temperatures. The fibre optics can be operated with certificated Matrix sensors, with an optical wave length from 500nm to 900nm. The fibre optics must not be buckled or laid with a small radius. Buckled or bad laid fibre optics results to a strong decrease of performance. Avoid performance decreasing and failures caused by wear, by a functional mounting of the fibre optics.

Maintenance

The fibre optics are maintenance-free. Protect the fibre optics against pollution. If they are contaminated, clean with alcohol. Do not use aggressive solvents. Equipment must only be repaired or serviced by the manufacturer.

Safety Informations

When installing and operating, it is necessary to take into consideration the relevant international and other national regulations. EN 60079-14, ATEX 118a, single directive 1999/92/EC.

Standards met:

EN 60079-0:2012 + A11:2013, EN 60079-28:2015, EN 60079-31:2014,

EN 13463-1:2009, EN 60529:2014

ATEX directive: 2014/34/EU, Machine directive: 2006/42/EC,

RoHS directive: 2011/65/EU

General Notes, disposal

We reserve the right to modify our equipment. Our equipment is designed such way, that it has the least possible adverse effect on the environment. No longer usable or irreparable units must be disposed of in accordance with local waste disposal regulations.

EU-Declaration of Conformity

IECEx certification No. BVS 14.0108X.

<http://iecex.iec.ch/iecex/iecexweb.nsf/0/FE79714C0BAEF6F5C1257D7E0044F6A9?opendocument>

ATEX certification: Certification No. BVS 10 ATEX E 130 X, DEKRA EXAM GmbH, Zertifizierungsstelle, Carl-Beyling-Haus, Dinendahlstrasse 9, D-44809 Bochum, Ident-No. 0158.

ATEX certification of quality type production of Ex devices in accordance to the ATEX directive 2014/34/EU, CE 0158. Certification No: BVS 18 ATEX ZQS / E118, QAR No. DE/BVS/QAR13.0004/04. The conformity of the devices with the EC standards and directives and the EC-type examination certificate and the observation of the Quality Safety System ISO 9001:2015 with the ATEX module "Production", declares:

Pablo Ledergerber, Matrix Elektronik AG